

## 紡織實驗室

**Test Report** 報告號碼: TX10126B /2021 /LE 日期: 2021 年 01 月 11 日 頁數: 三 之 一 頁

恆大股份有限公司  
台北市林森北路 372 號力霸大樓 10 樓

我們依照客戶的要求, 根據申請者送交的樣品進行測試如下:

樣品敘述 : 醫用口罩  
顏色 : 天空藍  
成份 : 紡結/熔噴/紡結不織布  
最終產品 : 醫用口罩  
生產 / 供應商 : 恆大股份有限公司  
原產國家 : 台灣  
申請者 : 恆大股份有限公司  
收件日期 : 2021 年 01 月 05 日  
測試週期 : 2021 年 01 月 05 日至 2021 年 01 月 11 日

測試要求 : 依申請者要求檢測中華民國國家標準 CNS 15290 L1036 紡織品安全規範(一般要求)部份項目:  
1) 鉛  
2) 鎘

測試結果 : 請見下面說明

台灣檢驗科技股份有限公司

張家豪

張家豪  
副理



"Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, and such sample(s) are retained for 1 months only." This test report cannot be reproduced, except in full, without prior written permission of the Company.

"This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>.

Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law."

# 紡織實驗室

## Test Report 報告號碼: TX10126B /2021 /LE 日期: 2021 年 01 月 11 日 頁數: 三 之 二 頁

測試結果:

### 1. 鉛

測試方法: 依中華民國國家標準 CNS 4797-2 Z 7066-2, 使用感應耦合電漿-光學放射光譜儀

	結果 (毫克/公斤)	CNS 15290 上限值 (毫克/公斤)
	1	
鉛 (Pb)	未偵測出	90

樣品敘述:

1. 天空藍醫用口罩

備註: 報告極限值: 1 毫克/公斤

### 2. 鎘

測試方法: 依中華民國國家標準 CNS 4797-2 Z 7066-2, 使用感應耦合電漿-光學放射光譜儀

	結果 (毫克/公斤)		CNS 15290 上限值 (毫克/公斤)
	1	2	
鎘 (Cd)	未偵測出	未偵測出	不得使用

樣品敘述:

1. 天空藍醫用口罩-塑膠條

2. 天空藍醫用口罩-金屬條

備註: 報告極限值: 1 毫克/公斤

以上測試由 SGS 相關實驗室執行

"Unless otherwise stated the results shown in this test report refer only to the sample(s) tested, and such sample(s) are retained for 1 months only." This test report cannot be reproduced, except in full, without prior written permission of the Company.

"This document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at <http://www.sgs.com.tw/Terms-and-Conditions> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at <http://www.sgs.com.tw/Terms-and-Conditions>. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law."

樣品照片



\*\*\* 結 束 \*\*\*